


MICROTOME

<ul style="list-style-type: none"> ➤ Make /Model: Leica RM2125 RTS ➤ Specifications <ul style="list-style-type: none"> ○ Section thickness setting range: 0.5 - 60 μm ○ Vertical specimen stroke: 59 mm ○ Specimen retraction: ON/OFF approx. 20 μm ○ Specimen orientation: XY - $\pm 8^\circ$ ○ Trimming thickness: 10 μm, 50 μm 	
---	--

Testing Charges				
Testing Name	Industrial	Academic Institutes	Students and Faculty of IISU	
Sectioning and embedding of tissues	Rs.1,000/-	Rs. 500/-	Rs. 200/-	Per Sample (Max. Sample: 5)
Charges are on per hour basis.				